

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 04/12/2023

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:42
INTERNAL STANDARD	VERIFIED	08:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:43
BLANK TEST	0.00	08:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:44
BLANK TEST	0.00	08:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:45
BLANK TEST	0.00	08:46

PO Steve Bunio

