

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 04/26/2023

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:11
INTERNAL STANDARD	VERIFIED	14:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:12
BLANK TEST	0.00	14:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:13
BLANK TEST	0.00	14:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	14:14
BLANK TEST	0.00	14:15

*Handwritten signature: A. Steen, 4/26/23*

