

SOLUTION CHANGE

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 05/01/2023

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:52
INTERNAL STANDARD	VERIFIED	12:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:53
BLANK TEST	0.00	12:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:54
BLANK TEST	0.00	12:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:55
BLANK TEST	0.00	12:56

*Handwritten signature: Steven T Bunio*

