

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 05/08/2023

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:27
INTERNAL STANDARD	VERIFIED	08:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:28
BLANK TEST	0.00	08:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:29
BLANK TEST	0.00	08:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:30
BLANK TEST	0.00	08:31

*Donnie J Long*  
844

