

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/22/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:41
INTERNAL STANDARD	VERIFIED	07:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:42
BLANK TEST	0.00	07:42
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	07:43
BLANK TEST	0.00	07:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:44
BLANK TEST	0.00	07:45

Donnie J Long
1344

