

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 05/29/2023

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:56
INTERNAL STANDARD	VERIFIED	12:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	12:57
BLANK TEST	0.00	12:58
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	12:58
BLANK TEST	0.00	12:59
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	12:59
BLANK TEST	0.00	13:00



PO- [Signature] 2714