

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 06/07/2023

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:08
INTERNAL STANDARD	VERIFIED	07:08
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:08
BLANK TEST	0.00	07:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:10
BLANK TEST	0.00	07:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:11
BLANK TEST	0.00	07:12

*PO J. Fish #407*

