

SUPERVISOR TEST


WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 06/16/2023

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:51
INTERNAL STANDARD	VERIFIED	15:51
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	15:52
BLANK TEST	0.00	15:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:53
BLANK TEST	0.00	15:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:54
BLANK TEST	0.00	15:55



*Signature: C. B. Geisler*  
#224

SUPERVISOR TEST

-----  
WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606  
-----

Date: 06/16/2023

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:41
INTERNAL STANDARD	VERIFIED	15:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	15:42

Communication Error 1

