

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 06/29/2023

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:08
INTERNAL STANDARD	VERIFIED	08:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:08
BLANK TEST	0.00	08:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:09
BLANK TEST	0.00	08:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	08:11
BLANK TEST	0.00	08:11

