

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 08/08/2023

OPERATOR NAME:  
STEVEN T BUNIO  
Permit #: 44983  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:47
INTERNAL STANDARD	VERIFIED	10:47
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:48
BLANK TEST	0.00	10:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:49
BLANK TEST	0.00	10:50
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:50
BLANK TEST	0.00	10:51

*ps Steve Bunio #44983*

