

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 08/15/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:38
INTERNAL STANDARD	VERIFIED	07:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:38
BLANK TEST	0.00	07:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:39
BLANK TEST	0.00	07:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:41
BLANK TEST	0.00	07:42

Donnie J Long
844

