

SUPERVISOR TEST


WARREN CO.
 SHERIFF'S OFFICE
 DATAMASTER dmt: 106606

Date: 08/22/2023

OPERATOR NAME:
 JOHN D BATEHOLTS
 Permit #: 43007
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:23
INTERNAL STANDARD	VERIFIED	08:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:23
BLANK TEST	0.00	08:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:25
BLANK TEST	0.00	08:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:26
BLANK TEST	0.00	08:27



Handwritten signature and number: P. Bateholts #714