

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 106606

Date: 09/08/2023

OPERATOR NAME:  
CHARLES B GEISLER  
Permit #: 33757  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:29
INTERNAL STANDARD	VERIFIED	11:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:30
BLANK TEST	0.00	11:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:31
BLANK TEST	0.00	11:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	11:32
BLANK TEST	0.00	11:33

*Charles B. Geisler* #24

