

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/14/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:00
INTERNAL STANDARD	VERIFIED	07:00
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	07:01
BLANK TEST	0.00	07:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:02
BLANK TEST	0.00	07:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	07:03
BLANK TEST	0.00	07:04

