

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 10/29/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:53
INTERNAL STANDARD	VERIFIED	09:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:53
BLANK TEST	0.00	09:54
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:54
BLANK TEST	0.00	09:55
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	09:56
BLANK TEST	0.00	09:57

Donnie J Long 844

