

SUPERVISOR TEST




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmat: 106606

Date: 11/22/2023

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:06
INTERNAL STANDARD	VERIFIED	08:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:07
BLANK TEST	0.00	08:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:08
BLANK TEST	0.00	08:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:09
BLANK TEST	0.00	08:10



John D Bateholts - 714