

SOLUTION CHANGE




WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 106606

Date: 12/01/2023

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:46
INTERNAL STANDARD	VERIFIED	08:46
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	08:48
BLANK TEST	0.00	08:49
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	08:49
BLANK TEST	0.00	08:50



Handwritten signature