

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/02/2023

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:48
INTERNAL STANDARD	VERIFIED	00:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	00:48
BLANK TEST	0.00	00:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:50
BLANK TEST	0.00	00:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	00:51
BLANK TEST	0.00	00:52



Michael A Campbell #757