

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 01/23/2023

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:26
INTERNAL STANDARD	VERIFIED	14:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:26
BLANK TEST	0.00	14:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	14:27
BLANK TEST	0.00	14:28
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:29
BLANK TEST	0.00	14:29

*PO Fish #427*

