

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 01/30/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22390

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:57
INTERNAL STANDARD	VERIFIED	11:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:58
BLANK TEST	0.00	11:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:59
BLANK TEST	0.00	12:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:00
BLANK TEST	0.00	12:01

