

SOLUTION CHANGE

WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 02/01/2023

OPERATOR NAME:
 JOSHUA M FISH
 Permit #: 37493
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:24
INTERNAL STANDARD	VERIFIED	09:24
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:24
BLANK TEST	0.00	09:25
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:25
BLANK TEST	0.00	09:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:27
BLANK TEST	0.00	09:28

PO #127

