

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/15/2023

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:51
INTERNAL STANDARD	VERIFIED	09:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:51
BLANK TEST	0.00	09:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:52
BLANK TEST	0.00	09:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:53
BLANK TEST	0.00	09:54

PC #427

