

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 02/22/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:32
INTERNAL STANDARD	VERIFIED	08:32
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:33
BLANK TEST	0.00	08:34
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:34
BLANK TEST	0.00	08:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:35
BLANK TEST	0.00	08:36

