

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/08/2023

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 22420

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:49
INTERNAL STANDARD	VERIFIED	01:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:49
BLANK TEST	0.00	01:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:52
BLANK TEST	0.00	01:53

[Handwritten signature]
#820

