

SOLUTION CHANGE




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 03/31/2023

OPERATOR NAME:
MICHAEL A CAMPBELL
Permit #: 42422
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:52
INTERNAL STANDARD	VERIFIED	01:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:52
BLANK TEST	0.00	01:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:54
BLANK TEST	0.00	01:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	01:55
BLANK TEST	0.00	01:56



PO MLO #757