

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/07/2023

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:13
INTERNAL STANDARD	VERIFIED	10:13
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:13
BLANK TEST	0.00	10:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:14
BLANK TEST	0.00	10:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:16
BLANK TEST	0.00	10:17



Handwritten signature and number: 2774