

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 04/19/2023

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:05
INTERNAL STANDARD	VERIFIED	14:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:05
BLANK TEST	0.00	14:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:07
BLANK TEST	0.00	14:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	14:08
BLANK TEST	0.00	14:09

*Donnie J Long*  
844

