

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 04/26/2023

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:27
INTERNAL STANDARD	VERIFIED	12:27
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:28
BLANK TEST	0.00	12:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:29
BLANK TEST	0.00	12:30
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:30
BLANK TEST	0.00	12:31

Handwritten signatures and initials

