

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 05/08/2023

OPERATOR NAME:  
DONNIE J LONG  
Permit #: 47653  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:16
INTERNAL STANDARD	VERIFIED	10:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:17
BLANK TEST	0.00	10:18
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:18
BLANK TEST	0.00	10:19
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	10:19
BLANK TEST	0.00	10:20

*Donnie J Long*  
844

