

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/15/2023

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:52
INTERNAL STANDARD	VERIFIED	08:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:53
BLANK TEST	0.00	08:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:54
BLANK TEST	0.00	08:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:55
BLANK TEST	0.00	08:56

Handwritten signature

