

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 05/29/2023

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23030

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:20
INTERNAL STANDARD	VERIFIED	13:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:20
BLANK TEST	0.00	13:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	13:21
BLANK TEST	0.00	13:22
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:23
BLANK TEST	0.00	13:23



Per *SB* #714