

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/07/2023

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:49
INTERNAL STANDARD	VERIFIED	11:49
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:49
BLANK TEST	0.00	11:50
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:50
BLANK TEST	0.00	11:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	11:52
BLANK TEST	0.00	11:53

[Handwritten Signature]
#427

