

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 06/22/2023

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:00
INTERNAL STANDARD	VERIFIED	12:00
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:00
BLANK TEST	0.00	12:01
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:01
BLANK TEST	0.00	12:02
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:02
BLANK TEST	0.00	12:03

PO Joshua M Fish #427

