

SUPERVISOR TEST




WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/16/2023

OPERATOR NAME:
JOHN D BATEHOLTS
Permit #: 43007
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:44
INTERNAL STANDARD	VERIFIED	08:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:45
BLANK TEST	0.00	08:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:46
BLANK TEST	0.00	08:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	08:47
BLANK TEST	0.00	08:48



P.C. *[Signature]* 714