

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 07/23/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23090

Nominal = 0.10 g/210L

BLANK TEST	0.00	11:31
INTERNAL STANDARD	VERIFIED	11:31
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	11:31
BLANK TEST	0.00	11:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:32
BLANK TEST	0.00	11:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	11:33
BLANK TEST	0.00	11:34

Donnie J Long
844

