

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/01/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:54
INTERNAL STANDARD	VERIFIED	17:54
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:55
BLANK TEST	0.00	17:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.10	17:56
BLANK TEST	0.00	17:57
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:57
BLANK TEST	0.00	17:58

[Handwritten signature] *[Handwritten initials]*

