

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 08/08/2023

OPERATOR NAME:
STEVEN T BUNIO
Permit #: 44983
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:15
INTERNAL STANDARD	VERIFIED	09:15
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:15
BLANK TEST	0.00	09:16
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:17
BLANK TEST	0.00	09:17
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:18
BLANK TEST	0.00	09:19

PO Steven Bunio 44983

