

SUPERVISOR TEST




WARREN COUNTY
 SHERIFF'S OFFICE
 DATAMASTER dmt: 144506

Date: 08/22/2023

OPERATOR NAME:
 JOHN D BATEHOLTS
 Permit #: 43007
 Operator Agency/Dept: WARREN CO. SO
 Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	08:42
INTERNAL STANDARD	VERIFIED	08:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:43
BLANK TEST	0.00	08:43
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:44
BLANK TEST	0.00	08:45
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	08:45
BLANK TEST	0.00	08:46



P. D. Bateholts 6714