

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/08/2023

OPERATOR NAME:
CHARLES B GEISLER
Permit #: 33757
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	13:42
INTERNAL STANDARD	VERIFIED	13:42
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:43
BLANK TEST	0.00	13:44
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	13:44
BLANK TEST	0.00	13:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	13:45
BLANK TEST	0.00	13:46

Charles B. Geisler

