

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 09/22/2023

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23190

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:06
INTERNAL STANDARD	VERIFIED	15:06
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:06
BLANK TEST	0.00	15:07
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:07
BLANK TEST	0.00	15:08
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	15:08
BLANK TEST	0.00	15:09

PO [Signature]
#127

