

SUPERVISOR TEST

WARREN COUNTY  
SHERIFF'S OFFICE  
DATAMASTER dmt: 144506

Date: 10/08/2023

OPERATOR NAME:  
JOSHUA M FISH  
Permit #: 37493  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:51
INTERNAL STANDARD	VERIFIED	17:51
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:51
BLANK TEST	0.00	17:52
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:52
BLANK TEST	0.00	17:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:54
BLANK TEST	0.00	17:54

*DO NOT SIGN*

