

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/14/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:33
INTERNAL STANDARD	VERIFIED	12:33
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:34
BLANK TEST	0.00	12:35
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:35
BLANK TEST	0.00	12:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:36
BLANK TEST	0.00	12:37

Donnie J Long

