

SUPERVISOR TEST

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 10/29/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	17:39
INTERNAL STANDARD	VERIFIED	17:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:39
BLANK TEST	0.00	17:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	17:41
BLANK TEST	0.00	17:41
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	17:42
BLANK TEST	0.00	17:43

