

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 11/01/2023

OPERATOR NAME:
DONNIE J LONG
Permit #: 47653
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23270

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:36
INTERNAL STANDARD	VERIFIED	10:36
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:36
BLANK TEST	0.00	10:37
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:38
BLANK TEST	0.00	10:39
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:39
BLANK TEST	0.00	10:40

Don J Long 844

