

SOLUTION CHANGE

WARREN COUNTY
SHERIFF'S OFFICE
DATAMASTER dmt: 144506

Date: 12/01/2023

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:53
INTERNAL STANDARD	VERIFIED	09:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:54
BLANK TEST	0.00	09:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:55
BLANK TEST	0.00	09:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	09:56
BLANK TEST	0.00	09:57

