

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/08/2024

OPERATOR NAME:
SEAN T SMITH
Permit #: 46227
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:46
INTERNAL STANDARD	VERIFIED	10:46
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:46
BLANK TEST	0.00	10:47
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:48
BLANK TEST	0.00	10:48
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:49
BLANK TEST	0.00	10:50

