

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/15/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:53
INTERNAL STANDARD	VERIFIED	10:53
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:54
BLANK TEST	0.00	10:55
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:55
BLANK TEST	0.00	10:56
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:56
BLANK TEST	0.00	10:57

[Handwritten signature]
#570

