

SUPERVISOR TEST


WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 01/30/2024

OPERATOR NAME:
J RIEHL
Permit #: 47503
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23350

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:56
INTERNAL STANDARD	VERIFIED	10:57
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:57
BLANK TEST	0.00	10:58
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:58
BLANK TEST	0.00	10:59
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:59
BLANK TEST	0.00	11:00


#820

