

SUPERVISOR TEST

WARREN CO.  
SHERIFF'S OFFICE  
DATAMASTER dmt: 105208

Date: 02/08/2024

OPERATOR NAME:  
J RIEHL  
Permit #: 47503  
Operator Agency/Dept: WARREN CO. SO  
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:26
INTERNAL STANDARD	VERIFIED	10:26
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	10:26
BLANK TEST	0.00	10:27
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	10:27
BLANK TEST	0.00	10:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	10:28
BLANK TEST	0.00	10:29

*[Handwritten signature]*  
#820

