

SUPERVISOR TEST

WARREN CO.
SHERIFF'S OFFICE
DATAMASTER dmt: 105208

Date: 02/15/2024

OPERATOR NAME:
JOSHUA M FISH
Permit #: 37493
Operator Agency/Dept: WARREN CO. SO
Reference Standard #: 23430

Nominal = 0.10 g/210L

BLANK TEST	0.00	12:03
INTERNAL STANDARD	VERIFIED	12:03
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	12:03
BLANK TEST	0.00	12:04
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:04
BLANK TEST	0.00	12:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.09	12:06
BLANK TEST	0.00	12:06

PO J.M. Fish #427

